

C 687056 Telford & Wrekin Council:

STL Laboratory Number	Customer Sample Ref.	Comment
11707816	March Sample	VOC reporting limits raised due to nature of sample matrix. Monophenol sampled from 1 litre pet as no fixed subsample was available. Reporting limits raised for TPH due to sample matrix. Low TPH surrogate recovery at 43.2% due to interference from the sample matrix. SVOC reporting limits raised due to nature of sample matrix. Unable to report results for Demeton-s-methyl and Disulphoton due to AQC breaches.

Key

N/S - Not Scheduled

I/S - Insufficient Sample

To Follow - analysis incomplete (interim reports only)